

OXFORD LAB PRODUCTS

Benchmate MS1 Magnetic stirrer



Visit www.oxfordlp.com to see the complete product offering and learn more about our products.

Call us at 800-554-4920 to schedule additional product demos or learn about the best consumables for your equipment.

CONTENTS

1. Product Description.....	1
2. Intended Use.....	1
3. Safety Precautions.....	1
4. Features.....	3
5. Accessories.....	3
6. Technical Specifications.....	4
7. Installation & Operation.....	5
8. Warranty Statement.....	8
9. Product Disposal.....	10

1. PRODUCT DESCRIPTION

The Oxford MS1 magnetic stirrer is a slim & compact electromagnetic stirrer designed with modern magnetic coil technology. It provides high IP protection class (IP 65) without any moving parts attached to it. Setting of speed and time is simple & easy with a digital display. For better mixing, pulse mode is provided in which a stirring bar can rotate clockwise & counter clockwise for every 30 seconds.

2. INTENDED USE

The Oxford MS1 magnetic stirrer is designed for use in laboratories, pharmacies & schools for mixing liquids industry for mixing liquids.

3. SAFETY PRECAUTIONS

Read the following safety instructions before using the electromagnetic stirrer for your as well as equipment safety.

1. Do not place any steel or magnetic material on the base plate except the recommended magnetic stirring bar with beaker or flask in between. Doing this can affect the magnetism of the device.
2. Place the device on a flat, stable, clean, non-slippery and fire-proof surface.
3. The base plate must be cleaned before re-use. Use soft mild cloth to clean.
4. Do not try to mix liquid more that 800 ml with 25 mm magnetic stirring bar. It may lead to improper mixing.
5. Wear protective equipment in accordance with the hazardous category of the media to be processed. Otherwise there is a risk from:
 - ☞ Splashing and evaporation of liquids.
 - ☞ Ejection of parts.

☞ Release of toxic or combustible gases.

6. Do not lift or hold the device from the cable.
7. Do not use damaged beaker, flask, stirring bar or any other component for operation. It may affect the efficiency of device.
8. Reduce the speed:
 - ☞ If the medium splashes out of the vessel.
 - ☞ If the appliance is not running smoothly.
 - ☞ If the container moves on the base plate.

9. Wear protective gloves while cleaning the device.

10. Do not touch the base plate immediately after operation. The base plate may get warm after continuous usage.



11. Process pathogenic materials only in a closed vessel under a suitable extractor hood. Do not operate the appliance in explosive atmospheres, with hazardous substances or under water.

12. Only use liquid for mixing that will not react dangerously to the extra energy produced through processing. This also applies to any extra energy produced in other ways. For example: Through light irradiation, through surrounding temperature etc.



13. Abrasion of the dispersion equipment or the rotating accessories can get into medium you are working on.

14. Chemical reaction of PTFE can occur when in contact with molten or dissolved alkaline earth metals, as well as with fine partied powders of metals from the 2 and 3 group of the periodical table at temperatures above 300-400 °C. Only elementary fluorine, chlorine trifluoride and alkaline metals do attack PTFE. Halogen

hydrocarbons have a reversible swelling effect.

15. The temperature of the stirring media in the beaker and beaker itself should not be more than 45° C.



16. The equipment will not restart, both in case of power supply interruption and in the case of any fault or mechanical interruption. Care should be taken based on the sample used for mixing.

4. FEATURES

1. Slim profile with a unique flat surface design.
2. Wear-free drive without any moving parts.
3. Mixing capacity of 800 mL with 25mm magnetic stirring bar.
4. Pulse function of 30 seconds for better mixing.
5. Variable time setting from 1 minute to 99 minute & infinite mode.
6. Digital display for accurate setting of speed & time.
7. Variable speed setting from 15 to 1500 RPM.
8. Water spill resistant design
9. Manufactured to IP 65 standard.
10. Easy to use control panel with a user friendly interface.
11. Anti - Skid pads.

5. ACCESSORIES

Stirrer Bar (PTFE coated, 25 x 8mm high strength stir bar)





6. TECHNICAL SPECIFICATIONS

Stirring quantity max. per stirring position (H2O)	800 ml
Stirred quantity (Max) (Tested in H2O)	500 ml at 1500 RPM with 25mm stirring bar.
Speed range	15 - 1500 rpm
Time range	1 to 99 minute & infinite mode
Digital Display	Yes
Recommended bar length	25 mm
Set-up plate material	Stainless Steel
Set-up plate dimensions	Ø94 mm
Dimensions (L x W x H)	156 x 110 x 13.2 mm
Weight	0.4 - 0.5 Kg
Permissible ambient temperature	5 - 40°C
Protection class according to DIN EN 60529	IP 65
Input voltage	100 - 230 VAC, 50/60 Hz
Output Voltage	9VDC, 1A
Power Consumption	5W
Auto direction change/Pulse Mode	Yes
Auto direction change time interval	30 seconds

7. INSTALLATION & OPERATION

This electromagnetic stirrer is provided in a packaging box. Open the box, remove the packaging and gently place the device on the firm and leveled surface. Take care while unpacking and do not pull the AC power adaptor as it is attached with the device. The user manual should be kept near the device. Please keep all packaging in safe storage for at least two years for warranty purpose.

Below table shows the function of user interface buttons.

Buttons	Buttons Name	Function
	ON/OFF	Press button to turn ON and turn OFF the unit.
	Pulse & Time	Single press button to activate pulse mode. LED blinks to indicate. Long press button to activate time mode. "t" will show on display.
	Increment	Press button to increase speed/time value.
	Decrement	Press button to decrease speed/time value.

1. Connect the AC power adaptor to the main power supply.
2. Fill the beaker, flask or vessel with the medium or liquids to be processed then place the magnetic stirring bar in the beaker, flask or vessel.
3. Place the beaker, flask or vessel on the center of the base plate.
4. Turn ON the electromagnetic stirrer by pressing ON/OFF button.
5. Set the speed at which the medium or liquids needs to be processed. Minimum speed is 15 RPM and maximum speed is 1500 RPM.
6. Set the Time if required. Minimum time is 1 minute & Maximum time is 99 minutes. By default time is always infinite.

SPEED SETTING

- Press increment (+) & decrement (-) button to modify & set speed value.
- Single press increment (+) or decrement (-) button to increase or decrease the speed value by 50 RPM on each press.
- Continuous press increment (+) or decrement (-) button to increase or decrease the speed value at faster pace.
- Speed value is seen in the digital display and are one set after 3 blinks.
- At any point of time, user can modify the speed using increment (+) or decrement (-) button.
- On every restart, the stirrer will start with the last run speed.

TIME SETTING

The default timer is set as infinite when stirrer starts. In order to set time,

- Long press the “PULSE/TIME” button, to activate time mode. By default the timer will be infinite and will be seen as “t ∞” in display.
- Once activated, single press increment (+) or decrement (-) button to increase or decrease time value by 1 minute on each press.
- Once activated, continuous press increment (+) or decrement (-) button to increase or decrease time value as faster pace.
- After modifying, wait for 3 blinks to set time. Timer is countdown timer and it will take effect once it reaches set RPM.
- Speed can be modified at any point after you have set the time. Speed and time are both independent factors, hence the modified speed will not affect the set time. The set speed and time will be displayed alternatively when speed reaches to the set RPM.
- When timer goes to zero (0), the stirring stops. The display will show “t 00” & “0000” indicating time for stirring is over.
- Press any button except the power button to start the stirring operation with last run speed and time.

Note: Timer will be set to default (infinite) once the magnetic stirrer is restarted.

PULSE MODE

In pulse mode, the stirring direction will be reversed every 30 seconds. In order to select the pulse mode single press pulse button.

First set the required speed & then single press the pulse button to start stirring process in pulse mode. LED adjacent to pulse button will blink to indicate activation of the pulse mode.

You can also set a time for Pulse Mode. To do so, set speed and also time, then single press the pulse button to start stirring in pulse mode. Pulse stirring mode ends when time goes to zero (0).

Press the ON/OFF button again to stop the ongoing operation as well as to switch OFF the electromagnetic stirrer.

After every restart (ON/OFF), stirrer will start operating at last run speed & infinite time.

NOTE: In case of slippage, first turn OFF and turn ON again. If slippage still occurs, then reduce the speed for efficient operation.

8. WARRANTY STATEMENT

This products is warranted to be free from defects in material and workmanship for a period of two (2) years from date of purchase. Your product will be duly repaired upon prompt notification in compliance with the following conditions :

- This warranty is valid only if the product is used for its intended purpose and within the guidelines specified in this instruction manual. This warranty does not cover damage caused by accident, neglect, misuse, improper service, natural forces or other causes not arising from defects in original material or workmanship. This warranty does not cover any incidental or consequential damages, commercial loss or any other damages from the use of this product.
- The warranty is invalidated by any non-factory modification, which will immediately terminate all liabilities on us for the products or damages caused by its use. The customer shall be responsible for the product or use of products as well as any supervision required for safety. If requested the products must be returned to the distributor in well packed and insured manner and all shipping charges must be paid.

Note: Some states do not allow limitation on the length of implied warranties or the

exclusion or limitation of incidental or consequential damages. This warranty gives you specific legal rights. This warranty is given expressly in lieu of all other warranties, expressed or implied.

- The purchaser agrees that there is no warranty of merchantability or of fitness for any intended purpose and that there are no other remedies or warranties, expressed or implied, which extend beyond the description on the face of the agreement. This warranty is only applicable to the original purchaser.

For your reference, make a note of the serial number, date of purchase and supplier here.	
Serial No.:	Purchase Date:
Supplier:	

9. PRODUCT DISPOSAL

By observing proper disposal of the equipment, you will help conserve natural and environmental resources along with protecting human health.

This equipment is marked with the crossed out wheeled bin symbol to indicate that this equipment must not be disposed of with unsorted waste.

User is responsible for correctly disposing of the equipment by handing it over to an authorized facility for separate collection and recycling. Also, it's your responsibility to decontaminate the equipment in case of biological, chemical and/or radiological contamination, so as to protect from health hazards the persons involved in the disposal and recycling of the equipment.



Please contact your local dealer from whom you originally purchased this equipment to get more information about where you can drop off your waste equipment.

OXFORD **LAB PRODUCTS**

Visit www.oxfordlp.com to see the complete product offering and learn more about our products.

Call us at 800-554-4920 to schedule additional product demos or learn about the best consumables for your equipment.

Send warranty card information to:

Oxford Lab Products
10360 Sorrento Valley Road
San Diego, CA 92121

Thank You for choosing Oxford Lab Products for your research.