

# OXFORD LAB PRODUCTS

## Benchmate VM-M Vortex Mixer Mini



Visit [www.oxfordlp.com](http://www.oxfordlp.com) to see the complete product offering and learn more about our products.

Call us at 800-554-4920 to schedule additional product demos or learn about the best consumables for your equipment.



# CONTENTS

1. Introduction.....	1
2. Intended Use.....	1
3. Safety Precautions.....	1
4. Features.....	2
5. Standard Part Listing.....	2
6. Operation.....	3
7. Technical Specification.....	3
8. Maintenance & Cleaning.....	5
9. Troubleshooting.....	6
10. Warranty Statement.....	6
11. Product Disposal.....	7

# 1. INTRODUCTION

The Oxford VM-M mini vortex mixer is available in two variants: Fixed speed and Variable speed. It comes with a powerful brushless DC motor and delivers speed up to 4500 RPM. With an orbital motion of 4.5 mm it provides thorough mixing and it can accommodate plastic tubes & vials up to 30 mm diameter.

# 2. INTENDED USE

The Oxford VM-M mini vortex mixer is used in laboratories for the general mixing of applications in individual test tubes for up to 30 mm diameter.

# 3. SAFETY PRECAUTIONS

△ **Read the following safety instructions before using the Oxford VM-M mini vortex mixer for your safety & the safety of the equipment.**

1. Gently take the mini vortex mixer out of the packaging box and place it on a flat, stable, clean, non - slippery surface.
2. Do not lift or hold the device from the power cable.
3. This mini vortex mixer is to be used to mix material that the user knows will not react dangerously to any extra energy produced from mixing.
4. The mini vortex mixer should be used for the specified application only. It must not be operated in a hazardous or flammable environment and must not be used to mix explosive or highly reactive material.
5. Users should always use covered test tubes. Overfilling of liquid may cause the medium to splash out.
6. **Caution:** The equipment will restart, both in case of power interruption and in case of any fault or mechanical interruption. Care


should be taken based on sample used for mixing.

7. Always use tubes up to 30 mm diameter for best result and product safety.

## 4. FEATURES


- Touch Operation – can handle plastic tubes of up to 30 mm diameter
- Extremely small footprint
- 4.5mm orbital motion
- Suction cups for proper grip
- Die cast base for stable operation
- Low voltage power supply
- Suspension design for longer switch life
- Variable speed for gentle to vigorous mixing (Variable model)

## 5. STANDARD PART LISTING



## 6. OPERATION

Operation starts automatically when the soft rubber cup is pressed



down using the test tube (as shown in image) and stops once the test tube is removed. In the variable mini vortex mixer, rotate the speed knob to increase or decrease the speed between 2500 to 4500 RPM. Rotate clockwise to increase the speed and counter clockwise to decrease the speed

## 7. TECHNICAL SPECIFICATION


Specification	Variable	Fix
Type of movement	Orbital	
Shaker diameter	4.5 mm	
Tube diameter	Up to 30 mm	
Permissible shaking weight (Include attachment)	100 g	
Maximum Speed	4500 RPM	
Variable Speed	Yes	No
Speed Accuracy	± 5 %	
Operation Type	Touch	
Dimensions (L x W x H)	100 x 100 x 70 mm	
Weight	0.55 Kg	
Noise Level	<50 dB	
Permissible ambient temperature	5 - 40 °C	

Specification	Variable	Fix
Protection class according to DIN EN 60529	IP 43	
Permissible relative moisture	<80%	
Air Pressure	80 to 106 kPa	
Input Voltage	100 - 240 V, 50 / 60 Hz	
Output Voltage	9 V , 1.5A	
Power Consumption	14 W	
Rated input of adaptor	0.3A	

## 8. MAINTENANCE & CLEANING

1. While cleaning, the product must not be plugged in to the power supply.
2. The entire unit should be wiped down periodically with a soft cloth. Excessive amount of liquid should be avoided. Liquid should not come in contact with the motor.
3. After cleaning, ensure that all parts are dried before re-use.
4. Do not use the unit if it shows signs of mechanical or electrical damage. Do not use the equipment if the rubber cup or suction cup is damaged. In such case, see warranty & contact manufacturing if applicable
5. The brushless motor in the mini vortex mixer doesn't require routine maintenance. Any required service should be performed by authorized, qualified personnel only.

# 9. TROUBLESHOOTING



Small bottom screw at the base of device pointed by an arrow.

Vortex Action	Problem	Solution
Partially Working (stops in between)	Small bottom screw might have loosened. Need to adjust	Tighten the small bottom screw (1/2 or 1 turns max)
Not working at all. (Completely stopped)	Small bottom screw might have loosened. Need to tighten	Tighten the small bottom screw (1 or 2 turns max)
Continuously working (run without press)	Small Bottom screw might have tightened. Need to loosen.	Loosen the small bottom screw (1 or 2 turns max)

**Note:** If after implementation of the required solution does not produce normal vortex action, contact service representative.

# 10. WARRANTY STATEMENT

This products is warranted to be free from defects in material and workmanship for a period of two (2) years from date of purchase. Your product will be duly repaired upon prompt notification in compliance with the following conditions :

- This warranty is valid only if the product is used for its intended purpose and within the guidelines specified in this instruction manual. This warranty does not cover damage caused by accident, neglect, misuse, improper service, natural forces or other causes not arising from defects in original material or workmanship. This warranty does not cover any incidental or consequential damages, commercial loss or any other damages from the use of this product.
- The warranty is invalidated by any non-factory modification, which will immediately terminate all liabilities on us for the products or



damages caused by its use. The customer shall be responsible for the product or use of products as well as any supervision required for safety. If requested the products must be returned to the distributor in well packed and insured manner and all shipping charges must be paid.

**Note:** Some states do not allow limitation on the length of implied warranties or the exclusion or limitation of incidental or consequential damages. This warranty gives you specific legal rights. This warranty is given expressly in lieu of all other warranties, expressed or implied.

- The purchaser agrees that there is no warranty of merchantability or of fitness for any intended purpose and that there are no other remedies or warranties, expressed or implied, which extend beyond the description on the face of the agreement. This warranty is only applicable to the original purchaser.

Products received without proper authorization will not be processed for warranty or service . All items returned for service should be sent with postage prepaid in the original packaging or other suitable packaging, padded to avoid damage. We will not be responsible for damage incurred by improper packaging.

**Note:** This warranty is valid only if the warranty is registered with the supplier within 30 days from the date of purchase.

## 11. PRODUCT DISPOSAL

By observing proper disposal of the equipment, you will help conserve natural and environmental resources along with protecting human health.

This equipment is marked with the crossed out wheeled bin symbol to indicate that this equipment must not be disposed of with unsorted waste.

User is responsible for correctly disposing of the equipment by handing

it over to an authorized facility for separate collection and recycling. Also, it's your responsibility to decontaminate the equipment in case of biological, chemical and/or radiological contamination, so as to protect from health hazards the persons involved in the disposal and recycling of the equipment.



Please contact your local dealer from whom you originally purchased this equipment to get more information about where you can drop off your waste equipment.



# **OXFORD**

The logo for Oxford Lab Products features the word "OXFORD" in a bold, black, sans-serif font. The letter "O" is replaced by a stylized white drop shape with a black outline. Below "OXFORD" is the text "LAB PRODUCTS" in a smaller, bold, black, sans-serif font.

## **LAB PRODUCTS**

**Visit [www.oxfordlp.com](http://www.oxfordlp.com) to see the complete product offering and learn more about our products.**

**Call us at 800-554-4920 to schedule additional product demos or learn about the best consumables for your equipment.**

**Send warranty card information to:**

Oxford Lab Products  
10360 Sorrento Valley Road  
San Diego, CA 92121

**Thank You for choosing Oxford Lab Products for your research.**